

Certificate Of Analysis

WhatmanTM

Product Information

Product Number: 10538018

Product Name: CF12 58X58CM 100PK

Lot Number/Serial Number: 17458191

Date of Manufacture: 9-Oct-2021

Paper Lot: W201

Paper Description: 100% pure cotton fibers, with no additives

Quality Control

Analytical Data	Mean	CLSI Requirements
Mean pH	7.0	5.7-7.5 pH
Mean Serum Uptake (μl / 1/8" disc)	N/A	1.37-1.71 μl
Mean Blood absorption time (sec /100 μl)	8.1	5-30 sec
Mean Spot Diameter (mm)	16.2	15-17 mm

Note: The Mean Serum Uptake result is provided by the Centers for Disease Control and Prevention (CDC).

Conformance & Quality systems statement

This is to certify that this product conforms to Cytiva specifications.

Manufacture site certified to ISO 9001:2015 and ISO 13485: 2016 and those products tested in accordance with documented quality procedures and approved as a result of meeting the required specification.

Electronic signature

This document has been electronically produced and is valid without a signature.

Version AE



www.cytiva.com
Global Life Sciences Solutions Operations UK Ltd.

Amersham Place Little Chalfont

Buckinghamshire HP7 9NA UK